


ARTIFICIAL IRIS

TECHNICAL INFORMATION



Model	 Artificial Iris with Fiber	 Artificial Iris Fiber Free
Indication	For aniridia cases where suturing is indicated	For aniridia cases where suturing is not indicated
Material	Pigmented silicone elastomer with polymer fiber meshwork	Pigmented silicone elastomer without polymer fiber meshwork
Specifics	<p>The design is based on the labelled photo print-out (hard copy) serving as color target for the individual production and resembling the natural iris.</p> <p>The diameter of the implant can be individually trephined.</p>	
Availability	The delivery time is 4 to 8 weeks after receipt of the order and approved photos.	

For more information please visit www.artificial-iris.com